

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
 Data File : VD068134.D
 Acq On : 21 Jan 2021 15:52
 Operator : VA/SY
 Sample : M1198-02
 Misc : 4.60G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-037-IDWSO-20210120

Quant Time: Jan 22 00:10:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	390610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	691319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	593868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	251899	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	175665	47.01	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.02%
35) Dibromofluoromethane	7.920	113	173254	45.33	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.66%
50) Toluene-d8	10.344	98	686867	44.93	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.86%
62) 4-Bromofluorobenzene	12.632	95	229848	41.75	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.50%
Target Compounds						
					Qvalue	
5) Bromomethane	2.773	94	11724	3.08	ug/l #	79
10) Methyl Iodide	4.309	142	10570	5.23	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
Data File : VD068134.D
Acq On : 21 Jan 2021 15:52
Operator : VA/SY
Sample : M1198-02
Misc : 4.60G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-037-IDWSO-20210120

Quant Time: Jan 22 00:10:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

