

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
 Data File : VD068135.D
 Acq On : 21 Jan 2021 16:19
 Operator : VA/SY
 Sample : M1193-01
 Misc : 6.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-012121

Quant Time: Jan 22 00:10:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	540547	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	913148	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	884294	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	400760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	267793	51.79	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 103.58%			
35) Dibromofluoromethane	7.920	113	267094	52.91	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.82%			
50) Toluene-d8	10.344	98	1137999	56.35	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 112.70%			
62) 4-Bromofluorobenzene	12.632	95	420004	57.76	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 115.52%			
Target Compounds						Qvalue
52) Toluene	10.403	92	18593	1.11	ug/l	97
95) Naphthalene	15.402	128	530123	35.03	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
Data File : VD068135.D
Acq On : 21 Jan 2021 16:19
Operator : VA/SY
Sample : M1193-01
Misc : 6.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-012121

Quant Time: Jan 22 00:10:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

