

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
 Data File : VD068137.D
 Acq On : 21 Jan 2021 17:14
 Operator : VA/SY
 Sample : M1177-01
 Misc : 4.89G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PILE-1-(0.5-1.0)

Quant Time: Jan 22 00:10:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	577022	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	982941	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	920097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	400098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	271406	49.17	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 98.34%			
35) Dibromofluoromethane	7.920	113	277498	51.06	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.12%			
50) Toluene-d8	10.344	98	1221130	56.18	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 112.36%			
62) 4-Bromofluorobenzene	12.632	95	418454	53.46	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 106.92%			
Target Compounds						
16) Acetone	4.162	43	177685	180.28	ug/l	95
25) 2-Butanone	7.214	43	25428	17.13	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
Data File : VD068137.D
Acq On : 21 Jan 2021 17:14
Operator : VA/SY
Sample : M1177-01
Misc : 4.89G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PILE-1-(0.5-1.0)

Quant Time: Jan 22 00:10:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

