

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012423\
 Data File : VD075189.D
 Acq On : 24 Jan 2023 15:58
 Operator : KP/SY
 Sample : 01283-04
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-10

Quant Time: Jan 25 00:33:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	80898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	158907	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.586	117	147314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	61878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	42930	52.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.000%
35) Dibromofluoromethane	7.810	113	47405	49.369	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.740%
50) Toluene-d8	10.269	98	197406	52.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.575	95	61738	51.453	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.900%
Target Compounds						
16) Acetone	4.028	43	5122	23.525	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012423\
Data File : VD075189.D
Acq On : 24 Jan 2023 15:58
Operator : KP/SY
Sample : 01283-04
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-10

Quant Time: Jan 25 00:33:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

