

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
 Data File : VD068213.D
 Acq On : 28 Jan 2021 10:31
 Operator : VA/SY
 Sample : VD0128SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0128SBL01

Quant Time: Jan 29 00:05:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	352360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	620095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.650	117	550744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	245954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	149570	44.37	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.74%
35) Dibromofluoromethane	7.920	113	150345	43.85	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.70%
50) Toluene-d8	10.338	98	565348	41.23	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.46%
62) 4-Bromofluorobenzene	12.632	95	205721	41.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
Data File : VD068213.D
Acq On : 28 Jan 2021 10:31
Operator : VA/SY
Sample : VD0128SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0128SBL01

Quant Time: Jan 29 00:05:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

