

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
 Data File : VD068220.D
 Acq On : 28 Jan 2021 13:50
 Operator : VA/SY
 Sample : M1254-06
 Misc : 5.67G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CRT-5

Quant Time: Jan 29 00:08:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	484570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	854063	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	757735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.578	152	330268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	263699	56.89	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.78%
35) Dibromofluoromethane	7.920	113	252471	53.47	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.94%
50) Toluene-d8	10.343	98	1060841	56.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.34%
62) 4-Bromofluorobenzene	12.631	95	340201	50.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.04%
Target Compounds						
95) Naphthalene	15.396	128	89265	7.16	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
Data File : VD068220.D
Acq On : 28 Jan 2021 13:50
Operator : VA/SY
Sample : M1254-06
Misc : 5.67G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CRT-5

Quant Time: Jan 29 00:08:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

