

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
 Data File : VD068229.D
 Acq On : 28 Jan 2021 17:56
 Operator : VA/SY
 Sample : M1271-02
 Misc : 7.18G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-332

Quant Time: Jan 29 00:11:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	461033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	814404	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	704456	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	283442	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	257747	58.44	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.88%
35) Dibromofluoromethane	7.920	113	242857	53.94	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.88%
50) Toluene-d8	10.344	98	1009169	56.03	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.06%
62) 4-Bromofluorobenzene	12.632	95	309878	47.78	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.56%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
Data File : VD068229.D
Acq On : 28 Jan 2021 17:56
Operator : VA/SY
Sample : M1271-02
Misc : 7.18G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-332

Quant Time: Jan 29 00:11:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

