

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
 Data File : VD068231.D
 Acq On : 28 Jan 2021 18:50
 Operator : VA/SY
 Sample : M1269-01
 Misc : 6.53G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-012821

Quant Time: Jan 29 00:11:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	469570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	823387	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	722859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	299845	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	256907	57.19	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.38%
35) Dibromofluoromethane	7.920	113	243442	53.48	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.96%
50) Toluene-d8	10.344	98	1009060	55.42	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.84%
62) 4-Bromofluorobenzene	12.632	95	318389	48.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.12%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012821\
Data File : VD068231.D
Acq On : 28 Jan 2021 18:50
Operator : VA/SY
Sample : M1269-01
Misc : 6.53G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-012821

Quant Time: Jan 29 00:11:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

