

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020823\
 Data File : VD075275.D
 Acq On : 08 Feb 2023 16:49
 Operator : KP/SY
 Sample : 01445-09
 Misc : 5.51G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JPP-34-020323

Quant Time: Feb 09 05:42:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 04:26:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	63061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	123772	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	122471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	52216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	39698	53.686	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	107.380%	
35) Dibromofluoromethane	7.810	113	36253	41.772	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	83.540%	
50) Toluene-d8	10.269	98	152757	46.412	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	92.820%	
62) 4-Bromofluorobenzene	12.575	95	49152	45.290	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	90.580%	
Target Compounds						
16) Acetone	4.022	43	4296	21.972	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020823\
Data File : VD075275.D
Acq On : 08 Feb 2023 16:49
Operator : KP/SY
Sample : 01445-09
Misc : 5.51G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JPP-34-020323

Quant Time: Feb 09 05:42:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 04:26:58 2023
Response via : Initial Calibration

