



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021023\
 Data File : VD075304.D
 Acq On : 10 Feb 2023 10:34
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
 Title : SW846 8260
 Last Update : Wed Feb 08 04:26:58 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	7077	PASS
75	95	30	60	57.7	18741	PASS
95	95	100	100	100.0	32485	PASS
96	95	5	9	7.0	2285	PASS
173	174	0.00	2	1.2	326	PASS
174	95	50	100	81.3	26403	PASS
175	174	5	9	7.9	2095	PASS
176	174	95	101	97.0	25600	PASS
177	176	5	9	7.2	1834	PASS