

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021323\
 Data File : VD075321.D
 Acq On : 13 Feb 2023 12:07
 Operator : KP/SY
 Sample : 01528-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-03-021023

Quant Time: Feb 14 01:18:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 04:26:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	43340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	75480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	78034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	38158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	24630	48.465	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.940%
35) Dibromofluoromethane	7.805	113	24979	47.196	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.400%
50) Toluene-d8	10.269	98	93443	46.555	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.100%
62) 4-Bromofluorobenzene	12.575	95	34519	52.157	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021323\
Data File : VD075321.D
Acq On : 13 Feb 2023 12:07
Operator : KP/SY
Sample : 01528-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-03-021023

Quant Time: Feb 14 01:18:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 04:26:58 2023
Response via : Initial Calibration

