

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021423\
 Data File : VD075345.D
 Acq On : 14 Feb 2023 16:06
 Operator : KP/SY
 Sample : 01541-02
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ASH

Quant Time: Feb 15 03:35:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 04:26:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	58987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	106089	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	108820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	51394	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	31113	44.982	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.960%
35) Dibromofluoromethane	7.804	113	34022	45.735	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.480%
50) Toluene-d8	10.269	98	133665	47.380	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.760%
62) 4-Bromofluorobenzene	12.575	95	46191	49.656	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.320%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021423\
Data File : VD075345.D
Acq On : 14 Feb 2023 16:06
Operator : KP/SY
Sample : 01541-02
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ASH

Quant Time: Feb 15 03:35:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 04:26:58 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1109
Delta R.T. 0.000 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06




Instrument :
MSVOA_D
ClientSampleId :
ASH


Tgt Ion:168 Resp: 58987
Ion Ratio Lower Upper
168 100
99 56.8 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 44.982 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.006 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06

Tgt Ion: 65 Resp: 31113
Ion Ratio Lower Upper
65 100
67 49.0 0.0 105.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD075345.D

Acq: 14 Feb 2023 16:06

Instrument :

MSVOA_D

ClientSampleId :

ASH





Tgt Ion:114 Resp: 106089

Ion Ratio Lower Upper

114 100

63 20.1 0.0 40.2

88 14.9 0.0 32.2



#35

Dibromofluoromethane

Concen: 45.735 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.006 min

Lab File: VD075345.D

Acq: 14 Feb 2023 16:06




Tgt Ion:113 Resp: 34022


Ion Ratio Lower Upper

113 100

111 100.0 80.1 120.1

192 18.0 15.4 23.0









#50
Toluene-d8
Concen: 47.380 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06




Instrument :
MSVOA_D
ClientSampleId :
ASH


Tgt Ion: 98 Resp: 133665
Ion Ratio Lower Upper
98 100
100 64.3 52.7 79.1



#62
4-Bromofluorobenzene
Concen: 49.656 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06

Tgt Ion: 95 Resp: 46191
Ion Ratio Lower Upper
95 100
174 83.6 0.0 161.2
176 79.0 0.0 160.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1
Delta R.T. -0.006 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06

Instrument :
MSVOA_D
ClientSampleId :
ASH

Tgt Ion:117 Resp: 108820
Ion Ratio Lower Upper
117 100
82 56.0 39.6 59.4
119 30.6 24.6 37.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. -0.000 min
Lab File: VD075345.D
Acq: 14 Feb 2023 16:06

Tgt Ion:152 Resp: 51394
Ion Ratio Lower Upper
152 100
115 56.9 28.8 86.4
150 156.2 0.0 348.6

