

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030119\  
 Data File : VD060909.D  
 Acq On : 1 Mar 2019 15:12  
 Operator : VA/AP  
 Sample : VSTDIC005  
 Misc : 5.00g/5ml/MSVOA D/SOIL  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Quant Time: Mar 02 01:09:08 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D030119S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 01 16:47:52 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.03  | 168  | 735308   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.20  | 114  | 1262764  | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 12.36 | 117  | 1051468  | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 14.51 | 152  | 450312   | 50.00 | ug/l  | 0.00     |

|                             |        |     |          |      |        |      |
|-----------------------------|--------|-----|----------|------|--------|------|
| System Monitoring Compounds |        |     |          |      |        |      |
| 33) 1,2-Dichloroethane-d4   | 7.45   | 65  | 28256    | 5.17 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =    | 10.34% |      |
| 35) Dibromofluoromethane    | 6.91   | 113 | 49548    | 5.25 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =    | 10.50% |      |
| 50) Toluene-d8              | 10.40  | 98  | 121539   | 4.76 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =    | 9.52%  |      |
| 62) 4-Bromofluorobenzene    | 13.55  | 95  | 47316    | 4.76 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =    | 9.52%  |      |

| Target Compounds              |      |     |        |        | Qvalue    |
|-------------------------------|------|-----|--------|--------|-----------|
| 2) Dichlorodifluoromethane    | 1.58 | 85  | 37804  | 4.715  | ug/l 100  |
| 3) Chloromethane              | 1.76 | 50  | 32347  | 6.562  | ug/l 98   |
| 4) Vinyl Chloride             | 1.84 | 62  | 28275  | 6.172  | ug/l 95   |
| 5) Bromomethane               | 2.15 | 94  | 6631   | 7.736  | ug/l # 92 |
| 6) Chloroethane               | 2.26 | 64  | 7257   | 5.838  | ug/l 99   |
| 7) Trichlorofluoromethane     | 2.52 | 101 | 35585  | 4.792  | ug/l 92   |
| 8) Diethyl Ether              | 2.87 | 74  | 14003  | 10.965 | ug/l 99   |
| 9) 1,1,2-Trichlorotrifluoroet | 3.14 | 101 | 45318  | 9.924  | ug/l 92   |
| 10) Methyl Iodide             | 3.30 | 142 | 28805  | 2.209  | ug/l 90   |
| 11) Tert butyl alcohol        | 4.04 | 59  | 9816   | 23.653 | ug/l # 79 |
| 12) 1,1-Dichloroethene        | 3.12 | 96  | 38579  | 11.308 | ug/l 98   |
| 13) Acrolein                  | 3.03 | 56  | 12192  | 52.988 | ug/l 90   |
| 14) Allyl chloride            | 3.62 | 41  | 66909  | 6.007  | ug/l 94   |
| 15) Acrylonitrile             | 4.18 | 53  | 39538  | 30.511 | ug/l 96   |
| 16) Acetone                   | 3.21 | 43  | 47227  | 28.482 | ug/l # 86 |
| 17) Carbon Disulfide          | 3.37 | 76  | 137045 | 5.437  | ug/l 99   |
| 18) Methyl Acetate            | 3.64 | 43  | 22295  | 6.481  | ug/l 99   |
| 19) Methyl tert-butyl Ether   | 4.23 | 73  | 67236  | 4.701  | ug/l # 89 |
| 20) Methylene Chloride        | 3.80 | 84  | 76271  | 8.188  | ug/l 89   |
| 21) trans-1,2-Dichloroethene  | 4.21 | 96  | 43941  | 5.812  | ug/l 97   |
| 22) Diisopropyl ether         | 5.10 | 45  | 130827 | 5.988  | ug/l # 91 |
| 23) Vinyl Acetate             | 5.03 | 43  | 325065 | 26.292 | ug/l 98   |
| 24) 1,1-Dichloroethane        | 4.97 | 63  | 70797  | 5.932  | ug/l 96   |
| 25) 2-Butanone                | 6.06 | 43  | 52229  | 23.866 | ug/l 98   |
| 26) 2,2-Dichloropropane       | 6.01 | 77  | 55913  | 4.935  | ug/l 91   |
| 27) cis-1,2-Dichloroethene    | 6.03 | 96  | 49577  | 5.972  | ug/l 91   |
| 28) Bromochloromethane        | 6.42 | 49  | 31839  | 6.781  | ug/l 89   |
| 29) Tetrahydrofuran           | 6.44 | 42  | 29138  | 26.719 | ug/l 93   |
| 30) Chloroform                | 6.66 | 83  | 73240  | 5.187  | ug/l 98   |
| 31) Cyclohexane               | 6.93 | 56  | 63963  | 6.461  | ug/l # 83 |
| 32) 1,1,1-Trichloroethane     | 6.86 | 97  | 60344  | 5.490  | ug/l 96   |
| 36) 1,1-Dichloropropene       | 7.13 | 75  | 60149  | 4.553  | ug/l 95   |
| 37) Ethyl Acetate             | 6.17 | 43  | 29512  | 4.751  | ug/l 98   |
| 38) Carbon Tetrachloride      | 7.10 | 117 | 57621  | 4.700  | ug/l 92   |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030119\  
 Data File : VD060909.D  
 Acq On : 1 Mar 2019 15:12  
 Operator : VA/AP  
 Sample : VSTDIC005  
 Misc : 5.00g/5ml/MSVOA D/SOIL  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Quant Time: Mar 02 01:09:08 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D030119S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 01 16:47:52 2019  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|--------|--------|----------|
| 39) Methylcyclohexane          | 8.85  | 83   | 61097    | 4.350  | ug/l   | 85       |
| 40) Benzene                    | 7.45  | 78   | 146767   | 4.470  | ug/l   | 99       |
| 41) Methacrylonitrile          | 6.42  | 41   | 30782    | 9.174  | ug/l   | 94       |
| 42) 1,2-Dichloroethane         | 7.56  | 62   | 38951    | 4.362  | ug/l   | 100      |
| 43) Isopropyl Acetate          | 9.23  | 43   | 35493    | 4.034  | ug/l # | 90       |
| 44) Trichloroethene            | 8.53  | 130  | 44314    | 4.139  | ug/l   | 99       |
| 45) 1,2-Dichloropropane        | 8.93  | 63   | 34883    | 4.548  | ug/l   | 93       |
| 46) Dibromomethane             | 9.05  | 93   | 22121    | 4.014  | ug/l   | 87       |
| 47) Bromodichloromethane       | 9.37  | 83   | 48757    | 3.731  | ug/l   | 93       |
| 48) Methyl methacrylate        | 9.11  | 41   | 21157    | 4.177  | ug/l   | 88       |
| 49) 1,4-Dioxane                | 9.07  | 88   | 3876     | 73.046 | ug/l # | 63       |
| 51) 4-Methyl-2-Pentanone       | 10.29 | 43   | 116212   | 21.413 | ug/l   | 99       |
| 52) Toluene                    | 10.50 | 92   | 92466    | 4.862  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene      | 10.90 | 75   | 46870    | 3.885  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene    | 10.02 | 75   | 58102    | 4.180  | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane      | 11.18 | 97   | 30012    | 4.601  | ug/l   | 91       |
| 56) Ethyl methacrylate         | 11.03 | 69   | 27350    | 3.684  | ug/l   | 96       |
| 57) 1,3-Dichloropropane        | 11.40 | 76   | 40673    | 3.786  | ug/l   | 97       |
| 58) 2-Chloroethyl Vinyl ether  | 9.85  | 63   | 100899   | 28.370 | ug/l   | 96       |
| 59) 2-Hexanone                 | 11.52 | 43   | 81773    | 19.625 | ug/l   | 98       |
| 60) Dibromochloromethane       | 11.68 | 129  | 38687    | 4.127  | ug/l   | 96       |
| 61) 1,2-Dibromoethane          | 11.80 | 107  | 32161    | 4.355  | ug/l   | 96       |
| 64) Tetrachloroethene          | 11.25 | 164  | 36358    | 3.870  | ug/l   | 94       |
| 65) Chlorobenzene              | 12.39 | 112  | 110000   | 5.162  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane  | 12.51 | 131  | 39854    | 4.699  | ug/l   | 98       |
| 67) Ethyl Benzene              | 12.52 | 91   | 173543   | 4.784  | ug/l   | 91       |
| 68) m/p-Xylenes                | 12.65 | 106  | 148061   | 11.549 | ug/l   | 82       |
| 69) o-Xylene                   | 13.04 | 106  | 66951    | 5.422  | ug/l   | 86       |
| 70) Styrene                    | 13.07 | 104  | 108325   | 5.220  | ug/l   | 95       |
| 71) Bromoform                  | 13.22 | 173  | 23489    | 3.483  | ug/l   | 99       |
| 73) Isopropylbenzene           | 13.39 | 105  | 166233   | 6.309  | ug/l   | 99       |
| 74) N-amyl acetate             | 13.26 | 43   | 48616    | 5.338  | ug/l   | 95       |
| 75) 1,1,2,2-Tetrachloroethane  | 13.70 | 83   | 33152    | 5.109  | ug/l   | 96       |
| 76) 1,2,3-Trichloropropane     | 13.73 | 75   | 33750    | 5.037  | ug/l   | 98       |
| 77) Bromobenzene               | 13.67 | 156  | 43620    | 5.144  | ug/l   | 83       |
| 78) n-propylbenzene            | 13.77 | 91   | 214747   | 6.108  | ug/l   | 97       |
| 79) 2-Chlorotoluene            | 13.83 | 91   | 113935   | 5.666  | ug/l   | 95       |
| 80) 1,3,5-Trimethylbenzene     | 14.23 | 105  | 136845   | 6.319  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-buten | 13.47 | 75   | 9099     | 4.262  | ug/l   | 88       |
| 82) 4-Chlorotoluene            | 13.94 | 91   | 133643   | 6.058  | ug/l   | 92       |
| 83) tert-Butylbenzene          | 14.18 | 119  | 149309   | 5.971  | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene     | 14.23 | 105  | 136845   | 6.319  | ug/l   | 98       |
| 85) sec-Butylbenzene           | 14.36 | 105  | 174532   | 6.252  | ug/l   | 91       |
| 86) p-Isopropyltoluene         | 14.48 | 119  | 136491   | 5.822  | ug/l   | 95       |
| 87) 1,3-Dichlorobenzene        | 14.45 | 146  | 74897    | 4.829  | ug/l   | 95       |
| 88) 1,4-Dichlorobenzene        | 14.53 | 146  | 76010    | 5.074  | ug/l   | 96       |
| 89) n-Butylbenzene             | 14.79 | 91   | 149336   | 6.148  | ug/l   | 95       |
| 90) Hexachloroethane           | 15.01 | 117  | 40389    | 6.071  | ug/l   | 90       |
| 91) 1,2-Dichlorobenzene        | 14.80 | 146  | 71698    | 5.552  | ug/l   | 90       |
| 92) 1,2-Dibromo-3-Chloropropan | 15.37 | 75   | 3433     | 3.777  | ug/l   | 59       |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030119\  
Data File : VD060909.D  
Acq On : 1 Mar 2019 15:12  
Operator : VA/AP  
Sample : VSTDICC005  
Misc : 5.00g/5ml/MSVOA D/SOIL  
ALS Vial : 3 Sample Multiplier: 1


Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDICC005

Quant Time: Mar 02 01:09:08 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D030119S.M  
Quant Title : SW846 8260  
QLast Update : Fri Mar 01 16:47:52 2019  
Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 15.94 | 180  | 39383    | 4.309 | ug/l  | 96       |
| 94) Hexachlorobutadiene    | 16.03 | 225  | 23795    | 4.339 | ug/l  | 95       |
| 95) Naphthalene            | 16.12 | 128  | 55523    | 4.030 | ug/l  | 98       |
| 96) 1,2,3-Trichlorobenzene | 16.27 | 180  | 21539    | 3.065 | ug/l  | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.03 min Scan# 570

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion:168 Resp: 735308

Ion Ratio Lower Upper

168 100

99 57.2 38.2 57.4





Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC

7.03

Time-->



#2

Dichlorodifluoromethane

Concen: 4.715 ug/l

RT: 1.58 min Scan# 16

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 85 Resp: 37804

Ion Ratio Lower Upper

85 100

87 32.3 16.1 48.2





Abundance Ion 84.95 (84.65 to 85.65): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.58

Time-->









#3  
Chloromethane  
Concen: 6.562 ug/l  
RT: 1.76 min Scan# 34  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 50 Resp: 32347  
Ion Ratio Lower Upper  
50 100  
52 32.2 26.8 40.2



#4  
Vinyl Chloride  
Concen: 6.172 ug/l  
RT: 1.84 min Scan# 43  
Delta R.T. -0.02 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 62 Resp: 28275  
Ion Ratio Lower Upper  
62 100  
64 35.9 26.3 39.5









#5  
Bromomethane  
Concen: 7.736 ug/l  
RT: 2.15 min Scan# 74  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 94 Resp: 6631  
Ion Ratio Lower Upper  
94 100  
96 109.2 81.0 121.4  
93 26.9 0.0 0.0#



#6  
Chloroethane  
Concen: 5.838 ug/l  
RT: 2.26 min Scan# 85  
Delta R.T. -0.02 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 64 Resp: 7257  
Ion Ratio Lower Upper  
64 100  
66 32.8 25.8 38.6






#7




Trichlorofluoromethane  
 Concen: 4.792 ug/l  
 RT: 2.52 min Scan# 112  
 Delta R.T. -0.02 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Tot Ion:101 Resp: 35585  
 Ion Ratio Lower Upper  
 101 100  
 103 66.9 48.6 72.8




Abundance Ion 100.85 (100.55 to 101.55): V  
 Ion 102.85 (102.55 to 103.55): V





#8


Diethyl Ether  
 Concen: 10.965 ug/l  
 RT: 2.87 min Scan# 147  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion: 74 Resp: 14003  
 Ion Ratio Lower Upper  
 74 100  
 45 128.6 64.8 194.4



Abundance Ion 74.00 (73.70 to 74.70): VDC  
 Ion 45.05 (44.75 to 45.75): VDC





#9

1,1,2-Trichlorotrifluoroethane

Concen: 9.924 ug/l

RT: 3.14 min Scan# 175

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion:101 Resp: 45318

Ion Ratio Lower Upper

101 100

85 41.7 34.7 52.1




151 67.2 61.2 91.8



Abundance Ion 100.85 (100.55 to 101.55): V

Ion 84.95 (84.65 to 85.65): V

Ion 150.90 (150.60 to 151.60): V



#10

Methyl Iodide

Concen: 2.209 ug/l

RT: 3.30 min Scan# 191

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion:142 Resp: 28805

Ion Ratio Lower Upper

142 100

127 50.7 34.2 51.4



141 14.7 11.4 17.2




Abundance Ion 141.85 (141.55 to 142.55): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.85 (140.55 to 141.55): V






#11




Tert butyl alcohol  
 Concen: 23.653 ug/l  
 RT: 4.04 min Scan# 266  
 Delta R.T. -0.02 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Tot Ion: 59 Resp: 9816  
 Ion Ratio Lower Upper  
 59 100  
 57 0.0 5.8 8.8#




Abundance Ion 58.95 (58.65 to 59.65): VDC  
 Ion 56.95 (56.65 to 57.65): VDC





#12


1,1-Dichloroethene  
 Concen: 11.308 ug/l  
 RT: 3.12 min Scan# 173  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion: 96 Resp: 38579  
 Ion Ratio Lower Upper  
 96 100  
 61 138.4 112.4 168.6  
 98 61.8 51.0 76.6



Abundance Ion 95.90 (95.60 to 96.60): VDC  
 Ion 60.95 (60.65 to 61.65): VDC  
 Ion 97.85 (97.55 to 98.55): VDC





#13

Acrolein

Concen: 52.988 ug/l

RT: 3.03 min Scan# 163

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 56 Resp: 12192

Ion Ratio Lower Upper




56 100

55 62.4 56.8 85.2



Abundance Ion 55.95 (55.65 to 56.65): VDC

Ion 55.00 (54.70 to 55.70): VDC



#14

Allyl chloride

Concen: 6.007 ug/l

RT: 3.62 min Scan# 223

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 41 Resp: 66909

Ion Ratio Lower Upper

41 100

39 65.1 55.7 83.5



76 35.4 31.3 46.9




Abundance Ion 41.05 (40.75 to 41.75): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 75.95 (75.65 to 76.65): VDC





#15

Acrylonitrile

Concen: 30.511 ug/l

RT: 4.18 min Scan# 280

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 53 Resp: 39538

Ion Ratio Lower Upper

53 100

52 80.7 66.2 99.2




51 33.6 31.0 46.6



Abundance Ion 53.00 (52.70 to 53.70): VDC

Ion 52.00 (51.70 to 52.70): VDC

Ion 51.00 (50.70 to 51.70): VDC



#16

Acetone

Concen: 28.482 ug/l

RT: 3.21 min Scan# 182

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 43 Resp: 47227

Ion Ratio Lower Upper



43 100


58 29.6 18.4 27.6#



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC





#17

Carbon Disulfide

Concen: 5.437 ug/l

RT: 3.37 min Scan# 198

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 76 Resp: 137045

Ion Ratio Lower Upper




76 100

78 8.6 7.3 10.9



Abundance Ion 75.85 (75.55 to 76.55): VDC

Ion 77.85 (77.55 to 78.55): VDC



#18

Methyl Acetate

Concen: 6.481 ug/l

RT: 3.64 min Scan# 225

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 43 Resp: 22295

Ion Ratio Lower Upper



43 100


74 22.7 18.6 28.0



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 74.00 (73.70 to 74.70): VDC









#19  
Methyl tert-butyl Ether  
Concen: 4.701 ug/l  
RT: 4.23 min Scan# 285  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 73 Resp: 67236  
Ion Ratio Lower Upper  
73 100  
57 22.7 14.4 21.6#



#20  
Methylene Chloride  
Concen: 8.188 ug/l  
RT: 3.80 min Scan# 242  
Delta R.T. -0.02 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 84 Resp: 76271  
Ion Ratio Lower Upper  
84 100  
49 124.2 84.7 127.1  
51 37.0 26.4 39.6  
86 60.7 50.4 75.6





#21

trans-1,2-Dichloroethene

Concen: 5.812 ug/l

RT: 4.21 min Scan# 283

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 96 Resp: 43941

Ion Ratio Lower Upper




96 100

61 128.6 100.1 150.1

98 64.4 49.9 74.9



Abundance Ion 95.90 (95.60 to 96.60): VDC  
Ion 60.95 (60.65 to 61.65): VDC  
Ion 97.85 (97.55 to 98.55): VDC



#22

Diisopropyl ether

Concen: 5.988 ug/l

RT: 5.10 min Scan# 374

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Tgt Ion: 45 Resp: 130827


Ion Ratio Lower Upper

45 100



43 55.1 49.6 74.4


87 16.4 17.0 25.4#

59 7.6 7.9 11.9#



Abundance Ion 45.05 (44.75 to 45.75): VDC  
Ion 43.05 (42.75 to 43.75): VDC  
Ion 87.00 (86.70 to 87.70): VDC  
Ion 58.95 (58.65 to 59.65): VDC





#23

Vinyl Acetate

Concen: 26.292 ug/l

RT: 5.03 min Scan# 367

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 43 Resp: 325065

Ion Ratio Lower Upper




43 100

86 7.4 6.5 9.7



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 86.00 (85.70 to 86.70): VDC



#24

1,1-Dichloroethane

Concen: 5.932 ug/l

RT: 4.97 min Scan# 361

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 63 Resp: 70797

Ion Ratio Lower Upper

63 100

98 7.8 2.9 8.6



100 3.3 1.8 5.3




Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 97.95 (97.65 to 98.65): VDC

Ion 99.95 (99.65 to 100.65): VDC





#25

2-Butanone

Concen: 23.866 ug/l

RT: 6.06 min Scan# 471

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 43 Resp: 52229

Ion Ratio Lower Upper

43 100

72 22.0 16.8 25.2



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 72.00 (71.70 to 72.70): VDC

6.06

15000



10000

5000

0

Time-->

5.95 6.00 6.05 6.10 6.15



#26

2,2-Dichloropropane

Concen: 4.935 ug/l

RT: 6.01 min Scan# 466

Delta R.T. -0.02 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 77 Resp: 55913

Ion Ratio Lower Upper

77 100

97 28.0 11.7 35.1



Abundance Ion 76.95 (76.65 to 77.65): VDC

Ion 96.85 (96.55 to 97.55): VDC

6.01

15000


10000


5000

0

Time-->

5.90 6.00 6.10









#27

cis-1,2-Dichloroethene  
Concen: 5.972 ug/l  
RT: 6.03 min Scan# 468  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005




Tot Ion: 96 Resp: 49577  
Ion Ratio Lower Upper  
96 100  
61 118.7 0.0 259.0  
98 60.3 0.0 132.4




#28

Bromochloromethane  
Concen: 6.781 ug/l  
RT: 6.42 min Scan# 508  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 49 Resp: 31839  
Ion Ratio Lower Upper  
49 100  
129 0.0 0.0 3.8  
130 77.1 69.8 104.8









#29  
Tetrahydrofuran  
Concen: 26.719 ug/l  
RT: 6.44 min Scan# 510  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 42 Resp: 29138  
Ion Ratio Lower Upper  
42 100  
72 40.5 36.2 54.2  
71 39.1 35.4 53.0



#30  
Chloroform  
Concen: 5.187 ug/l  
RT: 6.66 min Scan# 532  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 83 Resp: 73240  
Ion Ratio Lower Upper  
83 100  
85 66.3 51.5 77.3









#31  
Cyclohexane  
Concen: 6.461 ug/l  
RT: 6.93 min Scan# 560  
Delta R.T. -0.02 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 56 Resp: 63963  
Ion Ratio Lower Upper  
56 100  
69 27.2 25.3 37.9  
84 77.4 78.1 117.1#



#32  
1,1,1-Trichloroethane  
Concen: 5.490 ug/l  
RT: 6.86 min Scan# 553  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 97 Resp: 60344  
Ion Ratio Lower Upper  
97 100  
99 59.4 51.0 76.4  
61 41.1 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 5.490 ug/l

RT: 7.45 min Scan# 612

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 65 Resp: 28256

Ion Ratio Lower Upper




65 100

67 54.8 0.0 106.0



Abundance Ion 64.95 (64.65 to 65.65): VDC

Ion 66.90 (66.60 to 67.60): VDC



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.20 min Scan# 689

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 114 Resp: 1262764

Ion Ratio Lower Upper

114 100

63 16.4 0.0 34.8



88 16.2 0.0 39.0




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 62.95 (62.65 to 63.65): VDC

Ion 87.90 (87.60 to 88.60): VDC





#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 6.91 min Scan# 558

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion:113 Resp: 49548

Ion Ratio Lower Upper

113 100

111 97.4 81.9 122.9




192 14.7 16.9 25.3#



Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V



#36

1,1-Dichloropropene

Concen: 4.553 ug/l

RT: 7.13 min Scan# 580

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 75 Resp: 60149

Ion Ratio Lower Upper

75 100

110 37.4 16.8 50.3



77 30.4 25.8 38.8




Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 109.90 (109.60 to 110.60): V

Ion 76.95 (76.65 to 77.65): VDC









#37  
Ethyl Acetate  
Concen: 4.751 ug/l  
RT: 6.17 min Scan# 482  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleID :  
VSTDIC005


Tot Ion: 43 Resp: 29512  
Ion Ratio Lower Upper  
43 100  
61 16.8 12.5 18.7  
70 10.3 7.9 11.9



#38  
Carbon Tetrachloride  
Concen: 4.700 ug/l  
RT: 7.10 min Scan# 577  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 117 Resp: 57621  
Ion Ratio Lower Upper  
117 100  
119 87.8 77.2 115.8  
121 29.0 25.0 37.4





#39

Methvlcvclohexane

Concen: 4.350 ug/l

RT: 8.85 min Scan# 755

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 83 Resp: 61097

Ion Ratio Lower Upper




83 100

55 87.3 58.8 88.2

98 54.0 36.8 55.2



Abundance Ion 83.05 (82.75 to 83.75): VDC  
Ion 55.00 (54.70 to 55.70): VDC  
Ion 98.05 (97.75 to 98.75): VDC



#40

Benzene

Concen: 4.470 ug/l

RT: 7.45 min Scan# 612

Delta R.T. 0.00 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 78 Resp: 146767



Ion Ratio Lower Upper


78 100

77 23.3 18.9 28.3



Abundance Ion 77.95 (77.65 to 78.65): VDC  
Ion 76.95 (76.65 to 77.65): VDC





#41

Methacrylonitrile

Concen: 9.174 ug/l

RT: 6.42 min Scan# 508

Delta R.T. -0.02 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005

Tot Ion: 41 Resp: 30782


Ion Ratio Lower Upper

41 100

39 39.7 36.9 55.3

67 33.2 24.9 37.3

52 12.9 10.1 15.1






Abundance Ion 40.95 (40.65 to 41.65): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 67.00 (66.70 to 67.70): VDC

Ion 52.00 (51.70 to 52.70): VDC



#42

1,2-Dichloroethane

Concen: 4.362 ug/l

RT: 7.56 min Scan# 624

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 62 Resp: 38951

Ion Ratio Lower Upper



62 100


98 10.7 0.0 21.6



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 97.85 (97.55 to 98.55): VDC





#43

Isopropyl Acetate

Concen: 4.034 ug/l

RT: 9.23 min Scan# 793

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 43 Resp: 35493

Ion Ratio Lower Upper




43 100

61 37.3 25.4 38.2

87 0.0 0.3 0.5#



Abundance Ion 42.95 (42.65 to 43.65): VDC  
Ion 60.95 (60.65 to 61.65): VDC  
Ion 87.00 (86.70 to 87.70): VDC



#44

Trichloroethene

Concen: 4.139 ug/l

RT: 8.53 min Scan# 722

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion:130 Resp: 44314



Ion Ratio Lower Upper


130 100

95 90.7 0.0 179.6



Abundance Ion 129.80 (129.50 to 130.50): VDC  
Ion 94.90 (94.60 to 95.60): VDC





#45

1,2-Dichloropropane

Concen: 4.548 ug/l

RT: 8.93 min Scan# 763

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 63 Resp: 34883

Ion Ratio Lower Upper




63 100

65 34.8 24.6 36.8



Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 64.95 (64.65 to 65.65): VDC



#46

Dibromomethane

Concen: 4.014 ug/l

RT: 9.05 min Scan# 775

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 93 Resp: 22121

Ion Ratio Lower Upper

93 100

95 86.1 64.0 96.0



174 97.8 93.1 139.7




Abundance Ion 92.90 (92.60 to 93.60): VDC

Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): V





#47

Bromodichloromethane

Concen: 3.731 ug/l

RT: 9.37 min Scan# 807

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 83 Resp: 48757

Ion Ratio Lower Upper




83 100

85 69.3 50.8 76.2

127 8.3 7.3 10.9



Abundance Ion 82.85 (82.55 to 83.55): VDC  
Ion 84.95 (84.65 to 85.65): VDC  
Ion 126.80 (126.50 to 127.50): VDC



#48

Methyl methacrylate

Concen: 4.177 ug/l

RT: 9.11 min Scan# 781

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 41 Resp: 21157

Ion Ratio Lower Upper



41 100


69 66.7 61.4 92.2

39 44.1 42.1 63.1



Abundance Ion 40.95 (40.65 to 41.65): VDC  
Ion 69.00 (68.70 to 69.70): VDC  
Ion 38.95 (38.65 to 39.65): VDC









#49  
1,4-Dioxane  
Concen: 73.046 ug/l  
RT: 9.07 min Scan# 777  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005

Tot Ion: 88 Resp: 3876  
Ion Ratio Lower Upper  
88 100  
43 45.0 20.3 30.5#  
58 75.9 40.6 61.0#




Abundance Ion 88.00 (87.70 to 88.70): VDC  
Ion 43.05 (42.75 to 43.75): VDC  
Ion 57.95 (57.65 to 58.65): VDC






#50  
Toluene-d8  
Concen: 73.046 ug/l  
RT: 10.40 min Scan# 912  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 98 Resp: 121539  
Ion Ratio Lower Upper  
98 100  
100 68.4 55.3 82.9



Abundance Ion 98.05 (97.75 to 98.75): VDC  
Ion 100.05 (99.75 to 100.75): VDC





#51

4-Methyl-2-Pentanone

Concen: 21.413 ug/l

RT: 10.29 min Scan# 901

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 43 Resp: 116212

Ion Ratio Lower Upper




43 100

58 34.3 26.8 40.2



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC



#52

Toluene

Concen: 4.862 ug/l

RT: 10.50 min Scan# 922

Delta R.T. 0.00 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 92 Resp: 92466

Ion Ratio Lower Upper



92 100


91 167.0 132.1 198.1



Abundance Ion 92.00 (91.70 to 92.70): VDC

Ion 91.00 (90.70 to 91.70): VDC





#53

t-1,3-Dichloropropene

Concen: 3.885 ug/l

RT: 10.90 min Scan# 963

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 75 Resp: 46870

Ion Ratio Lower Upper




75 100

77 31.1 25.4 38.2



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC



#54

cis-1,3-Dichloropropene

Concen: 4.180 ug/l

RT: 10.02 min Scan# 874

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 75 Resp: 58102

Ion Ratio Lower Upper

75 100

77 35.3 26.1 39.1



39 46.3 36.6 54.8




Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC

Ion 38.95 (38.65 to 39.65): VDC





#55

1,1,2-Trichloroethane

Concen: 4.601 ug/l

RT: 11.18 min Scan# 991

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005

Tot Ion: 97 Resp: 30012


Ion Ratio Lower Upper

97 100

83 80.9 73.0 109.6

85 52.9 47.6 71.4

99 65.4 47.9 71.9






Abundance Ion 96.85 (96.55 to 97.55): VDC

Ion 82.95 (82.65 to 83.65): VDC

Ion 84.95 (84.65 to 85.65): VDC

Ion 98.85 (98.55 to 99.55): VDC



#56

Ethyl methacrylate

Concen: 3.684 ug/l

RT: 11.03 min Scan# 976

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion: 69 Resp: 27350

Ion Ratio Lower Upper

69 100

41 74.5 58.2 87.2



39 34.7 30.8 46.2




Abundance Ion 69.00 (68.70 to 69.70): VDC

Ion 41.05 (40.75 to 41.75): VDC

Ion 38.95 (38.65 to 39.65): VDC









#57  
 1,3-Dichloropropane  
 Concen: 3.786 ug/l  
 RT: 11.40 min Scan# 1014  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005


Tot Ion: 76 Resp: 40673  
 Ion Ratio Lower Upper  
 76 100  
 78 33.2 25.2 37.8



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 28.370 ug/l  
 RT: 9.85 min Scan# 856  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion: 63 Resp: 100899  
 Ion Ratio Lower Upper  
 63 100  
 106 30.8 23.0 34.4









#59  
2-Hexanone  
Concen: 19.625 ug/l  
RT: 11.52 min Scan# 1026  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 43 Resp: 81773  
Ion Ratio Lower Upper  
43 100  
58 52.4 25.4 76.2



#60  
Dibromochloromethane  
Concen: 4.127 ug/l  
RT: 11.68 min Scan# 1042  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 129 Resp: 38687  
Ion Ratio Lower Upper  
129 100  
127 80.7 38.5 115.3









#61  
 1,2-Dibromoethane  
 Concen: 4.355 ug/l  
 RT: 11.80 min Scan# 1054  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005


Tot Ion:107 Resp: 32161  
 Ion Ratio Lower Upper  
 107 100  
 109 97.5 75.0 112.6



#62  
 4-Bromofluorobenzene  
 Concen: 4.355 ug/l  
 RT: 13.55 min Scan# 1232  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion: 95 Resp: 47316  
 Ion Ratio Lower Upper  
 95 100  
 174 86.6 0.0 207.2  
 176 86.0 0.0 198.4









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 12.36 min Scan# 1111  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005

Tot Ion:117 Resp: 1051468  
Ion Ratio Lower Upper  
117 100  
82 44.6 45.5 68.3#  
119 32.3 25.7 38.5




Abundance Ion 116.95 (116.65 to 117.65): V  
Ion 82.05 (81.75 to 82.75): V  
Ion 118.95 (118.65 to 119.65): V






#64  
Tetrachloroethene  
Concen: 3.870 ug/l  
RT: 11.25 min Scan# 998  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion:164 Resp: 36358  
Ion Ratio Lower Upper  
164 100  
166 125.9 101.9 152.9  
129 87.8 63.0 94.4  
131 85.6 63.1 94.7



Abundance Ion 163.85 (163.55 to 164.55): V  
Ion 165.85 (165.55 to 166.55): V  
Ion 128.80 (128.50 to 129.50): V  
Ion 130.80 (130.50 to 131.50): V









#65  
Chlorobenzene  
Concen: 5.162 ug/l  
RT: 12.39 min Scan# 1114  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005

Tot Ion:112 Resp: 110000  
Ion Ratio Lower Upper  
112 100  
114 31.2 26.4 39.6




Abundance Ion 111.90 (111.60 to 112.60): V  
Ion 114.00 (113.70 to 114.70): V






#66  
1,1,1,2-Tetrachloroethane  
Concen: 4.699 ug/l  
RT: 12.51 min Scan# 1126  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion:131 Resp: 39854  
Ion Ratio Lower Upper  
131 100  
133 95.0 48.5 145.6  
119 66.7 32.8 98.3



Abundance Ion 130.90 (130.60 to 131.60): V  
Ion 132.90 (132.60 to 133.60): V  
Ion 118.85 (118.55 to 119.55): V









#67  
Ethyl Benzene  
Concen: 4.784 ug/l  
RT: 12.52 min Scan# 1127  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion: 91 Resp: 173543  
Ion Ratio Lower Upper  
91 100  
106 32.4 22.2 33.2



#68  
m/p-Xylenes  
Concen: 11.549 ug/l  
RT: 12.65 min Scan# 1141  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion: 106 Resp: 148061  
Ion Ratio Lower Upper  
106 100  
91 189.1 174.2 261.2





#69

o-Xylene

Concen: 5.422 ug/l

RT: 13.04 min Scan# 1180

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion:106 Resp: 66951

Ion Ratio Lower Upper




106 100

91 210.2 116.9 350.7



Abundance Ion 106.05 (105.75 to 106.75): V

Ion 91.00 (90.70 to 91.70): V



#70

Styrene

Concen: 5.220 ug/l

RT: 13.07 min Scan# 1183

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12


Tgt Ion:104 Resp: 108325

Ion Ratio Lower Upper

104 100

78 37.3 34.6 51.8



103 45.5 36.3 54.5




Abundance Ion 104.05 (103.75 to 104.75): V

Ion 77.95 (77.65 to 78.65): V

Ion 103.05 (102.75 to 103.75): V









#71  
Bromoform  
Concen: 3.483 ug/l  
RT: 13.22 min Scan# 1199  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion:173 Resp: 23489  
Ion Ratio Lower Upper  
173 100  
175 49.0 24.6 73.8  
254 6.3 6.0 9.0



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 14.51 min Scan# 1330  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion:152 Resp: 450312  
Ion Ratio Lower Upper  
152 100  
115 55.3 24.4 73.2  
150 159.7 0.0 311.0





#73

Isopropylbenzene

Concen: 6.309 ug/l

RT: 13.39 min Scan# 1216

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion:105 Resp: 166233

Ion Ratio Lower Upper




105 100

120 24.1 12.2 36.6



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V



#74

N-ethyl acetate

Concen: 5.338 ug/l

RT: 13.26 min Scan# 1203

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Tgt Ion: 43 Resp: 48616


Ion Ratio Lower Upper

43 100

70 32.3 26.8 40.2

55 19.9 14.2 21.4

61 29.9 20.6 31.0






Abundance Ion 42.95 (42.65 to 43.65): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 60.95 (60.65 to 61.65): V





#75

1,1,2,2-Tetrachloroethane

Concen: 5.109 ug/l

RT: 13.70 min Scan# 1247

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 83 Resp: 33152

Ion Ratio Lower Upper




83 100

131 9.7 4.9 14.6

85 60.2 31.9 95.9



Abundance Ion 82.95 (82.65 to 83.65): VDC  
Ion 130.90 (130.60 to 131.60): VDC  
Ion 84.95 (84.65 to 85.65): VDC



#76

1,2,3-Trichloropropane

Concen: 5.037 ug/l

RT: 13.73 min Scan# 1250

Delta R.T. 0.00 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 75 Resp: 33750



Ion Ratio Lower Upper


75 100

77 31.6 16.3 48.8



Abundance Ion 75.00 (74.70 to 75.70): VDC  
Ion 76.95 (76.65 to 77.65): VDC





#77

Bromobenzene

Concen: 5.144 ug/l

RT: 13.67 min Scan# 1244

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion:156 Resp: 43620

Ion Ratio Lower Upper

156 100

77 106.9 67.4 202.2




158 91.5 49.2 147.6



Abundance Ion 155.90 (155.60 to 156.60): V

Ion 76.95 (76.65 to 77.65): V

Ion 157.85 (157.55 to 158.55): V



#78

n-propylbenzene

Concen: 6.108 ug/l

RT: 13.77 min Scan# 1254

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 91 Resp: 214747

Ion Ratio Lower Upper



91 100


120 21.6 10.1 30.3



Abundance Ion 91.00 (90.70 to 91.70): V

Ion 120.05 (119.75 to 120.75): V





#79

2-Chlorotoluene

Concen: 5.666 ug/l

RT: 13.83 min Scan# 1261

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :


VSTDIC005

Tot Ion: 91 Resp: 113935

Ion Ratio Lower Upper




91 100

126 37.8 17.4 52.2



Abundance Ion 91.00 (90.70 to 91.70): VD060912.D (-1258) (-)

Ion 125.95 (125.65 to 126.65): VD060912.D (-1258) (-)



#80

1,3,5-Trimethylbenzene

Concen: 6.319 ug/l

RT: 14.23 min Scan# 1301

Delta R.T. -0.01 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion:105 Resp: 136845

Ion Ratio Lower Upper



105 100


120 44.4 22.8 68.4



Abundance Ion 105.05 (104.75 to 105.75): VD060912.D (-1293) (-)

Ion 120.05 (119.75 to 120.75): VD060912.D (-1293) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 4.262 ug/l

RT: 13.47 min Scan# 1224

Delta R.T. 0.00 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005





Tot Ion: 75 Resp: 9099

Ion Ratio Lower Upper

75 100

53 64.6 60.2 90.4

89 33.9 32.6 49.0



#82

4-Chlorotoluene

Concen: 6.058 ug/l

RT: 13.94 min Scan# 1272

Delta R.T. 0.00 min

Lab File: VD060909.D




Acq: 1 Mar 2019 15:12


Tgt Ion: 91 Resp: 133643

Ion Ratio Lower Upper

91 100

126 38.1 16.9 50.7









#83  
tert-Butylbenzene  
Concen: 5.971 ug/l  
RT: 14.18 min Scan# 1296  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12




Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005


Tot Ion:119 Resp: 149309  
Ion Ratio Lower Upper  
119 100  
91 66.2 34.2 102.5  
134 22.2 12.9 38.7



#84  
1,2,4-Trimethylbenzene  
Concen: 6.319 ug/l  
RT: 14.23 min Scan# 1301  
Delta R.T. -0.01 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion:105 Resp: 136845  
Ion Ratio Lower Upper  
105 100  
120 44.4 22.8 68.4









#85  
 sec-Butylbenzene  
 Concen: 6.252 ug/l  
 RT: 14.36 min Scan# 1314  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Instrument :  
 MSVOA\_D  
 ClientSampleID :  
 VSTDIC005

Tot Ion:105 Resp: 174532  
 Ion Ratio Lower Upper  
 105 100  
 134 16.2 10.1 30.3




Abundance Ion 105.05 (104.75 to 105.75): V  
 Ion 134.10 (133.80 to 134.80): V






#86  
 p-Isopropyltoluene  
 Concen: 5.822 ug/l  
 RT: 14.48 min Scan# 1327  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion:119 Resp: 136491  
 Ion Ratio Lower Upper  
 119 100  
 134 23.9 12.9 38.7  
 91 21.6 12.6 37.8



Abundance Ion 119.05 (118.75 to 119.75): V  
 Ion 134.10 (133.80 to 134.80): V  
 Ion 91.00 (90.70 to 91.70): V









#87  
 1,3-Dichlorobenzene  
 Concen: 4.829 ug/l  
 RT: 14.45 min Scan# 1324  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Tot Ion:146 Resp: 74897  
 Ion Ratio Lower Upper  
 146 100  
 111 37.2 15.9 47.6  
 148 66.6 32.4 97.0




Abundance Ion 145.95 (145.65 to 146.65): V  
 Ion 110.90 (110.60 to 111.60): V  
 Ion 147.90 (147.60 to 148.60): V






#88  
 1,4-Dichlorobenzene  
 Concen: 5.074 ug/l  
 RT: 14.53 min Scan# 1332  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion:146 Resp: 76010  
 Ion Ratio Lower Upper  
 146 100  
 111 37.3 15.9 47.7  
 148 66.4 32.6 97.8



Abundance Ion 145.95 (145.65 to 146.65): V  
 Ion 110.90 (110.60 to 111.60): V  
 Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 6.148 ug/l

RT: 14.79 min Scan# 1358

Delta R.T. -0.01 min

Lab File: VD060909.D

Acq: 1 Mar 2019 15:12

Instrument :

MSVOA\_D

ClientSampleId :

VSTDIC005


Tot Ion: 91 Resp: 149336

Ion Ratio Lower Upper

91 100

92 59.9 30.4 91.2




134 21.6 14.1 42.2



Abundance Ion 91.00 (90.70 to 91.70): VDC

Ion 92.00 (91.70 to 92.70): VDC

Ion 134.10 (133.80 to 134.80): VDC



#90

Hexachloroethane

Concen: 6.071 ug/l

RT: 15.01 min Scan# 1380

Delta R.T. 0.00 min

Lab File: VD060909.D


Acq: 1 Mar 2019 15:12

Tgt Ion: 117 Resp: 40389

Ion Ratio Lower Upper



117 100


201 74.0 41.4 124.4



Abundance Ion 116.85 (116.55 to 117.55): V

Ion 200.75 (200.45 to 201.45): V









#91  
 1,2-Dichlorobenzene  
 Concen: 5.552 ug/l  
 RT: 14.80 min Scan# 1359  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12




Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005


Tot Ion:146 Resp: 71698  
 Ion Ratio Lower Upper  
 146 100  
 111 43.6 16.7 50.1  
 148 61.1 32.6 97.8



#92  
 1,2-Dibromo-3-Chloropropane  
 Concen: 3.777 ug/l  
 RT: 15.37 min Scan# 1417  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion: 75 Resp: 3433  
 Ion Ratio Lower Upper  
 75 100  
 155 87.0 58.6 175.8  
 157 84.7 75.1 225.3









#93  
 1,2,4-Trichlorobenzene  
 Concen: 4.309 ug/l  
 RT: 15.94 min Scan# 1475  
 Delta R.T. 0.00 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VSTDIC005

Tot Ion:180 Resp: 39383  
 Ion Ratio Lower Upper  
 180 100  
 182 100.2 47.5 142.6  
 145 29.1 14.3 42.9




Abundance Ion 179.85 (179.55 to 180.55): V  
 Ion 181.85 (181.55 to 182.55): V  
 Ion 144.95 (144.65 to 145.65): V






#94  
 Hexachlorobutadiene  
 Concen: 4.339 ug/l  
 RT: 16.03 min Scan# 1484  
 Delta R.T. -0.01 min  
 Lab File: VD060909.D  
 Acq: 1 Mar 2019 15:12

Tgt Ion:225 Resp: 23795  
 Ion Ratio Lower Upper  
 225 100  
 223 66.1 30.6 91.6  
 227 66.3 32.0 96.0



Abundance Ion 224.75 (224.45 to 225.45): V  
 Ion 222.75 (222.45 to 223.45): V  
 Ion 226.75 (226.45 to 227.45): V









#95  
Naphthalene  
Concen: 4.030 ug/l  
RT: 16.12 min Scan# 1493  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Instrument :  
MSVOA\_D  
ClientSampleId :  
VSTDIC005

Tot Ion:128 Resp: 55523  
Ion Ratio Lower Upper  
128 100  
127 12.6 9.7 14.5  
129 12.2 9.0 13.4



#96  
1,2,3-Trichlorobenzene  
Concen: 3.065 ug/l  
RT: 16.27 min Scan# 1508  
Delta R.T. 0.00 min  
Lab File: VD060909.D  
Acq: 1 Mar 2019 15:12

Tgt Ion:180 Resp: 21539  
Ion Ratio Lower Upper  
180 100  
182 96.5 47.9 143.8  
145 26.9 12.9 38.7

