

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030519\
 Data File : VD060946.D
 Acq On : 5 Mar 2019 14:02
 Operator : VA/AP
 Sample : K1714-02RE
 Misc : 5.97g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-A-2-5RE

Quant Time: Mar 06 04:32:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 02 01:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	725311	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	1211322	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.37	117	1008342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	441916	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	216482	36.10	ug/l	0.02
Spiked Amount 50.000			Recovery	=	72.20%	
35) Dibromofluoromethane	6.93	113	366963	37.79	ug/l	0.02
Spiked Amount 50.000			Recovery	=	75.58%	
50) Toluene-d8	10.42	98	858786	35.54	ug/l	0.02
Spiked Amount 50.000			Recovery	=	71.08%	
62) 4-Bromofluorobenzene	13.56	95	341742	37.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.84%	
Target Compounds						
16) Acetone	3.23	43	7177	3.585	ug/l	95
20) Methylene Chloride	3.83	84	92974	6.636	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD030519\
Data File : VD060946.D
Acq On : 5 Mar 2019 14:02
Operator : VA/AP
Sample : K1714-02RE
Misc : 5.97g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-A-2-5RE

Quant Time: Mar 06 04:32:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 02 01:28:28 2019
Response via : Initial Calibration

