

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
 Data File : VD068553.D
 Acq On : 08 Mar 2021 11:32
 Operator : VA/SY
 Sample : VD0308SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0308SBL01

Quant Time: Mar 09 00:08:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	390588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	662498	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	633625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	291473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	185088	47.43	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.86%
35) Dibromofluoromethane	7.920	113	202016	50.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.24%
50) Toluene-d8	10.338	98	832300	52.36	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.72%
62) 4-Bromofluorobenzene	12.632	95	271132	52.92	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.84%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
Data File : VD068553.D
Acq On : 08 Mar 2021 11:32
Operator : VA/SY
Sample : VD0308SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0308SBL01

Quant Time: Mar 09 00:08:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

