

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
 Data File : VD068561.D
 Acq On : 08 Mar 2021 15:42
 Operator : VA/SY
 Sample : M1603-03
 Misc : 10.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 33-NORTH

Quant Time: Mar 09 00:11:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	248950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	464144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	391094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	121024	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	157221	63.21	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 126.42%			
35) Dibromofluoromethane	7.914	113	92735	33.17	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 66.34%			
50) Toluene-d8	10.344	98	534107	47.96	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 95.92%			
62) 4-Bromofluorobenzene	12.632	95	147798	41.18	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 82.36%			
Target Compounds					Qvalue	
16) Acetone	4.162	43	37974	96.11	ug/l	96
17) Carbon Disulfide	4.409	76	22048	2.69	ug/l	96
25) 2-Butanone	7.215	43	11728	20.65	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
Data File : VD068561.D
Acq On : 08 Mar 2021 15:42
Operator : VA/SY
Sample : M1603-03
Misc : 10.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
33-NORTH

Quant Time: Mar 09 00:11:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

