

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
 Data File : VD068562.D
 Acq On : 08 Mar 2021 16:10
 Operator : VA/SY
 Sample : M1603-01
 Misc : 8.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 31-NORTH

Quant Time: Mar 09 00:11:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	309872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	555829	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	546017	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	252449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	164597	53.16	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 106.32%		
35) Dibromofluoromethane	7.920	113	169727	50.69	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 101.38%		
50) Toluene-d8	10.343	98	705468	52.90	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 105.80%		
62) 4-Bromofluorobenzene	12.632	95	233970	54.43	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 108.86%		
Target Compounds						Qvalue
16) Acetone	4.161	43	9379	19.07	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
Data File : VD068562.D
Acq On : 08 Mar 2021 16:10
Operator : VA/SY
Sample : M1603-01
Misc : 8.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
31-NORTH

Quant Time: Mar 09 00:11:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

