

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
 Data File : VD068565.D
 Acq On : 08 Mar 2021 17:34
 Operator : VA/SY
 Sample : M1597-01
 Misc : 6.93G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 AU-06-030821

Quant Time: Mar 09 00:12:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	298254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	534571	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	499454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	193590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	161377	54.15	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.30%
35) Dibromofluoromethane	7.920	113	169295	52.57	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.14%
50) Toluene-d8	10.344	98	663127	51.70	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.40%
62) 4-Bromofluorobenzene	12.632	95	201952	48.85	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030821\
Data File : VD068565.D
Acq On : 08 Mar 2021 17:34
Operator : VA/SY
Sample : M1597-01
Misc : 6.93G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
AU-06-030821

Quant Time: Mar 09 00:12:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

