



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031023\
Data File : VD075486.D
Acq On : 10 Mar 2023 09:21
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031023S.M
Title : SW846 8260
Last Update : Fri Mar 10 23:52:37 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	10605	PASS
75	95	30	60	48.7	33037	PASS
95	95	100	100	100.0	67867	PASS
96	95	5	9	7.0	4727	PASS
173	174	0.00	2	0.9	505	PASS
174	95	50	100	86.3	58555	PASS
175	174	5	9	7.3	4269	PASS
176	174	95	101	96.4	56440	PASS
177	176	5	9	6.2	3519	PASS