

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
 Data File : VD068697.D
 Acq On : 29 Mar 2021 14:34
 Operator : VA/SY
 Sample : VD0329SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0329SBL01

Quant Time: Mar 30 01:04:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	335057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	645835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	651447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	301532	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	181158	53.62	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.24%
35) Dibromofluoromethane	7.914	113	203763	50.71	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.42%
50) Toluene-d8	10.338	98	822038	53.06	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.12%
62) 4-Bromofluorobenzene	12.632	95	279489	55.64	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
Data File : VD068697.D
Acq On : 29 Mar 2021 14:34
Operator : VA/SY
Sample : VD0329SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0329SBL01

Quant Time: Mar 30 01:04:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

