

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
 Data File : VD068700.D
 Acq On : 29 Mar 2021 16:05
 Operator : VA/SY
 Sample : M1823-01
 Misc : 6.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MK-3-29-B1

Quant Time: Mar 30 01:05:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	328214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	637870	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	598054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	301338	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	188295	56.89	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 113.78%		
35) Dibromofluoromethane	7.914	113	205948	51.89	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.78%		
50) Toluene-d8	10.338	98	772670	50.50	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 101.00%		
62) 4-Bromofluorobenzene	12.632	95	301031	60.68	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 121.36%		
Target Compounds						
16) Acetone	4.162	43	3621	9.02	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
Data File : VD068700.D
Acq On : 29 Mar 2021 16:05
Operator : VA/SY
Sample : M1823-01
Misc : 6.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MK-3-29-B1

Quant Time: Mar 30 01:05:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

