

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
 Data File : VD068703.D
 Acq On : 29 Mar 2021 17:31
 Operator : VA/SY
 Sample : M1810-02
 Misc : 7.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 600486548-BIN-1

Quant Time: Mar 30 01:06:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	293841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	585233	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	535209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	273304	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	173999	58.72	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.44%
35) Dibromofluoromethane	7.914	113	196225	53.89	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.78%
50) Toluene-d8	10.338	98	736141	52.44	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.88%
62) 4-Bromofluorobenzene	12.632	95	264137	58.03	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.06%
Target Compounds						Qvalue
44) Trichloroethene	9.120	130	6879	1.63	ug/l	73
64) Tetrachloroethene	10.873	164	3548	1.02	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
Data File : VD068703.D
Acq On : 29 Mar 2021 17:31
Operator : VA/SY
Sample : M1810-02
Misc : 7.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
600486548-BIN-1

Quant Time: Mar 30 01:06:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

