

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
 Data File : VD068706.D
 Acq On : 29 Mar 2021 18:56
 Operator : VA/SY
 Sample : M1828-02
 Misc : 6.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB13(25-26)

Quant Time: Mar 30 01:07:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	275287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	538287	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	501963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	266433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	156194	56.27	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.54%
35) Dibromofluoromethane	7.920	113	173878	51.92	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.84%
50) Toluene-d8	10.338	98	686057	53.13	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.26%
62) 4-Bromofluorobenzene	12.632	95	252890	60.41	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.82%
Target Compounds						
					Qvalue	
16) Acetone	4.161	43	3441	10.22	ug/l #	70
17) Carbon Disulfide	4.403	76	8399	1.20	ug/l #	88
30) Chloroform	7.714	83	7142	1.33	ug/l	96
64) Tetrachloroethene	10.879	164	12388	3.81	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032921\
Data File : VD068706.D
Acq On : 29 Mar 2021 18:56
Operator : VA/SY
Sample : M1828-02
Misc : 6.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB13(25-26)

Quant Time: Mar 30 01:07:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

