

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
 Data File : VD068720.D
 Acq On : 30 Mar 2021 11:09
 Operator : VA/SY
 Sample : VD0330SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0330SBL01

Quant Time: Mar 31 00:59:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	560454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	988484	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	984530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	480727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	257741	45.61	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.22%
35) Dibromofluoromethane	7.914	113	301087	48.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.92%
50) Toluene-d8	10.338	98	1219160	51.42	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.84%
62) 4-Bromofluorobenzene	12.632	95	436837	56.82	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.64%
Target Compounds						
52) Toluene	10.402	92	18247	1.12	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
Data File : VD068720.D
Acq On : 30 Mar 2021 11:09
Operator : VA/SY
Sample : VD0330SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0330SBL01

Quant Time: Mar 31 00:59:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

