

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
 Data File : VD068723.D
 Acq On : 30 Mar 2021 12:41
 Operator : VA/SY
 Sample : M1828-01
 Misc : 6.18G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB13(5-6)

Quant Time: Mar 31 01:00:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	502607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	894295	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	905579	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	445732	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	252015	49.73	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.46%
35) Dibromofluoromethane	7.920	113	276302	49.66	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.32%
50) Toluene-d8	10.338	98	1097781	51.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.34%
62) 4-Bromofluorobenzene	12.632	95	410507	59.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.04%
Target Compounds						
16) Acetone	4.162	43	11018	17.92	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
Data File : VD068723.D
Acq On : 30 Mar 2021 12:41
Operator : VA/SY
Sample : M1828-01
Misc : 6.18G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB13(5-6)

Quant Time: Mar 31 01:00:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

