

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
 Data File : VD068724.D
 Acq On : 30 Mar 2021 13:10
 Operator : VA/SY
 Sample : M1828-04RE
 Misc : 7.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB14(5-6)RE

Quant Time: Mar 31 01:00:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	54393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	98728	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	115802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	59453	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	42245	77.02	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.04%
35) Dibromofluoromethane	7.914	113	34572	56.28	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.56%
50) Toluene-d8	10.338	98	122933	51.91	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.82%
62) 4-Bromofluorobenzene	12.632	95	51860	67.54	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	135.08%
Target Compounds						
					Qvalue	
16) Acetone	4.168	43	8552	128.52	ug/l #	88
40) Benzene	8.356	78	4113	1.60	ug/l #	53
52) Toluene	10.408	92	2237	1.37	ug/l	89
95) Naphthalene	15.390	128	17681	9.58	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033021\
Data File : VD068724.D
Acq On : 30 Mar 2021 13:10
Operator : VA/SY
Sample : M1828-04RE
Misc : 7.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB14(5-6)RE

Quant Time: Mar 31 01:00:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

