



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040723\
Data File : VD075595.D
Acq On : 07 Apr 2023 10:23
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031023S.M
Title : SW846 8260
Last Update : Fri Mar 10 23:52:37 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	9850	PASS
75	95	30	60	49.5	32312	PASS
95	95	100	100	100.0	65224	PASS
96	95	5	9	7.1	4610	PASS
173	174	0.00	2	1.0	606	PASS
174	95	50	100	90.5	59024	PASS
175	174	5	9	7.4	4358	PASS
176	174	95	101	98.1	57877	PASS
177	176	5	9	6.8	3935	PASS