

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041323\
 Data File : VD075661.D
 Acq On : 13 Apr 2023 15:31
 Operator : KP/SY
 Sample : VD0413SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0413SBL01

Quant Time: Apr 14 01:34:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	171882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	327849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	311045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	130421	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	101390	62.146	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.300%
35) Dibromofluoromethane	7.805	113	118248	53.625	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.240%
50) Toluene-d8	10.269	98	425331	48.166	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.340%
62) 4-Bromofluorobenzene	12.575	95	115640	49.269	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041323\
Data File : VD075661.D
Acq On : 13 Apr 2023 15:31
Operator : KP/SY
Sample : VD0413SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0413SBL01

Quant Time: Apr 14 01:34:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

