

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041323\
 Data File : VD075668.D
 Acq On : 13 Apr 2023 19:04
 Operator : KP/SY
 Sample : 02305-01
 Misc : 6.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-232

Quant Time: Apr 14 01:36:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	105021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	305425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	391616	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	193396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	79704	79.956	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	159.920%
35) Dibromofluoromethane	7.804	113	119377	58.111	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.220%
50) Toluene-d8	10.269	98	426945	51.898	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.800%
62) 4-Bromofluorobenzene	12.575	95	157978	72.248	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	144.500%#
Target Compounds						
						Qvalue
20) Methylene Chloride	4.793	84	18860	9.626	ug/l #	77
70) Styrene	12.134	104	8802	1.405	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041323\
Data File : VD075668.D
Acq On : 13 Apr 2023 19:04
Operator : KP/SY
Sample : 02305-01
Misc : 6.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-232

Quant Time: Apr 14 01:36:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

