

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041719\
 Data File : VD061846.D
 Acq On : 17 Apr 2019 11:58
 Operator : VA/AP
 Sample : VD0417SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0417SBL01

Quant Time: Apr 17 15:42:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	627250	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.22	114	978886	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.36	117	766473	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	14.51	152	361934	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	298296	56.74	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	113.48%	
35) Dibromofluoromethane	6.92	113	436227	58.99	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	117.98%	
50) Toluene-d8	10.40	98	1053939	60.15	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	120.30%	
62) 4-Bromofluorobenzene	13.55	95	379510	53.43	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	106.86%	
Target Compounds						
20) Methylene Chloride	3.82	84	52578	3.938	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041719\
Data File : VD061846.D
Acq On : 17 Apr 2019 11:58
Operator : VA/AP
Sample : VD0417SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0417SBL01

Quant Time: Apr 17 15:42:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

