

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD041719\
 Data File : VD061852.D
 Acq On : 17 Apr 2019 14:59
 Operator : VA/AP
 Sample : K2390-05
 Misc : 8.23g/5ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 200050

Quant Time: Apr 17 15:47:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 08:53:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	492261	50.00	ug/l	-0.06
34) 1,4-Difluorobenzene	8.20	114	711258	50.00	ug/l	-0.06
63) Chlorobenzene-d5	12.36	117	616318	50.00	ug/l	-0.04
72) 1,4-Dichlorobenzene-d4	14.51	152	233922	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	177154	42.94	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	85.88%	
35) Dibromofluoromethane	6.91	113	291713	54.29	ug/l	-0.06
Spiked Amount 50.000			Recovery	=	108.58%	
50) Toluene-d8	10.40	98	630947	49.55	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	13.55	95	242055	46.90	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	93.80%	
Target Compounds						
16) Acetone	3.21	43	41026	40.348	ug/l	94
20) Methylene Chloride	3.81	84	82792	12.985	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD041719\
Data File : VD061852.D
Acq On : 17 Apr 2019 14:59
Operator : VA/AP
Sample : K2390-05
Misc : 8.23g/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
200050

Quant Time: Apr 17 15:47:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 08:53:01 2019
Response via : Initial Calibration

