


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD041719\
Data File : VD061844.D
Acq On : 17 Apr 2019 9:41
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Title : SW846 8260
Last Update : Tue Apr 16 08:53:01 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	12935	PASS
75	95	30	60	40.4	29560	PASS
95	95	100	100	100.0	73093	PASS
96	95	5	9	6.4	4665	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.8	63442	PASS
175	174	5	9	6.4	4066	PASS
176	174	95	101	98.5	62490	PASS
177	176	5	9	6.0	3744	PASS