

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042222\
 Data File : VD072690.D
 Acq On : 22 Apr 2022 13:20
 Operator : VA/SY
 Sample : N2494-04
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-1D

Quant Time: Apr 22 23:44:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	351826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	628755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	631472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	287789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	233508	54.865	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.740%	
35) Dibromofluoromethane	7.909	113	218208	44.657	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 89.320%	
50) Toluene-d8	10.332	98	747769	44.025	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 88.060%	
62) 4-Bromofluorobenzene	12.620	95	278530	41.989	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.980%	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	14883	18.754	ug/l #	85
17) Carbon Disulfide	4.409	76	2123	3.914	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042222\
Data File : VD072690.D
Acq On : 22 Apr 2022 13:20
Operator : VA/SY
Sample : N2494-04
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-1D

Quant Time: Apr 22 23:44:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

