

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042222\
 Data File : VD072692.D
 Acq On : 22 Apr 2022 14:17
 Operator : VA/SY
 Sample : N2525-03
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-02-042122

Quant Time: Apr 22 23:45:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	403242	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	731706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	709932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	328259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	253823	52.034	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.060%
35) Dibromofluoromethane	7.909	113	238271	41.902	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.800%
50) Toluene-d8	10.332	98	876490	44.343	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.680%
62) 4-Bromofluorobenzene	12.620	95	311629	40.369	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.740%
Target Compounds						
16) Acetone	4.162	43	13107	14.410	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042222\
Data File : VD072692.D
Acq On : 22 Apr 2022 14:17
Operator : VA/SY
Sample : N2525-03
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-02-042122

Quant Time: Apr 22 23:45:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

