




Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042723\
 Data File : VD075865.D
 Acq On : 27 Apr 2023 17:07
 Operator : KP/SY
 Sample : 02525-20
 Misc : 6.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCM3

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 04/28/2023
 Supervised By :Semsettin Yesilyurt 05/01/2023

Quant Time: Apr 28 02:58:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 27 02:02:18 2023
 Response via : Initial Calibration



TIC: VD075865.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.310min (+ 0.082) 0.04 ug/L

response 270

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	56.40	50.74
51.00	97.70	83.33
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042723\
 Data File : VD075865.D
 Acq On : 27 Apr 2023 17:07
 Operator : KP/SY
 Sample : 02525-20
 Misc : 6.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCM3

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 04/28/2023
 Supervised By :Semsettin Yesilyurt 05/01/2023

Quant Time: Apr 28 02:58:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 27 02:02:18 2023
 Response via : Initial Calibration



TIC: VD075865.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.228min (-0.000) 20.91 ug/L m

response 145507

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	56.40	0.09#
51.00	97.70	0.15#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042723\
 Data File : VD075865.D
 Acq On : 27 Apr 2023 17:07
 Operator : KP/SY
 Sample : 02525-20
 Misc : 6.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCM3

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 04/28/2023
 Supervised By :Semsettin Yesilyurt 05/01/2023

Quant Time: Apr 28 02:58:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 27 02:02:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	651777	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	650237	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	299246	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	241186	16.868	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	67.480%		
7) Chloroethane-d5	2.799	69	221629	17.940	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	71.760%		
11) 1,1-Dichloroethene-d2	3.917	65	57991	17.592	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	70.360%		
21) 2-Butanone-d5	6.993	46	65804	53.013	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	106.020%		
24) Chloroform-d	7.569	84	329744	20.231	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.920%		
26) 1,2-Dichloroethane-d4	8.228	65	145507m	20.905	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.640%		
32) Benzene-d6	8.205	84	649098	17.127	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	68.520%		
36) 1,2-Dichloropropane-d6	9.210	67	200646	19.127	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	76.520%		
41) Toluene-d8	10.269	98	597595	16.843	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	67.360%		
43) trans-1,3-Dichloroprop...	10.528	79	80172	18.490	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	73.960%		
47) 2-Hexanone-d5	10.881	63	58220	46.094	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	92.180%		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	179810	22.400	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	89.600%		
66) 1,2-Dichlorobenzene-d4	13.816	152	214670	20.281	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	81.120%		
Target Compounds					Qvalue	
13) Acetone	4.022	43	22394	21.082	ug/L	64
16) Methylene chloride	4.805	84	24475	1.851	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042723\
Data File : VD075865.D
Acq On : 27 Apr 2023 17:07
Operator : KP/SY
Sample : 02525-20
Misc : 6.00G/10.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCCM3

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 04/28/2023
Supervised By :Semsettin Yesilyurt 05/01/2023

Quant Time: Apr 28 02:58:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 27 02:02:18 2023
Response via : Initial Calibration

