



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042723\
 Data File : VD075853.D
 Acq On : 27 Apr 2023 10:37
 Operator : KP/SY
 Sample : BFB120
 Misc : 5.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB120

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Title : SFAM01.0
 Last Update : Tue May 02 02:31:38 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	11751	PASS
75	95	30	80	48.0	34675	PASS
95	95	100	100	100.0	72243	PASS
96	95	5	9	7.1	5140	PASS
173	174	0.00	2	0.7	464	PASS
174	95	50	120	91.7	66235	PASS
175	174	5	9	7.3	4811	PASS
176	174	95	105	96.6	63987	PASS
177	176	5	10	6.7	4313	PASS

Scan 1860 (12.528 min): VD075853.D\data.ms
BFB120

m/z	Abundance
40.00	1168.0
44.00	532.0
206.90	208.0

Instrument :
MSVOA_D
ClientSampleId :
BFB120

m/z	Abundance
36.00	301.0
37.00	2564.0
38.00	2554.0
39.00	1192.0
40.00	1063.0
44.00	966.0
45.00	729.0
47.10	941.0
48.00	428.0
49.00	3202.0
50.10	12557.0
51.00	3504.0
52.10	248.0
54.00	166.0
55.00	221.0
56.00	1151.0
57.00	2526.0
58.00	163.0
60.00	718.0
61.00	4052.0
62.00	4025.0
63.00	3071.0
63.90	585.0
65.10	661.0
66.80	212.0
68.00	8254.0
69.00	7335.0
70.00	581.0
71.90	279.0
73.00	3874.0
74.10	14339.0
75.00	38144.0
76.10	3301.0
77.20	510.0
78.00	385.0
79.00	4356.0
79.90	878.0
81.00	3960.0
82.00	778.0
86.00	358.0
87.00	4184.0
88.00	4549.0
91.00	320.0
92.00	2324.0
93.00	4341.0
94.10	10007.0
95.10	79680.0
96.00	5763.0
96.90	185.0
103.00	171.0
104.00	311.0
106.00	537.0
114.90	192.0
115.90	367.0
117.00	677.0
117.90	396.0
118.90	465.0
125.00	448.0
127.90	296.0
130.00	272.0
133.10	382.0
133.90	226.0
135.10	307.0
141.00	822.0
143.00	1271.0
144.80	505.0
157.00	183.0
171.00	305.0
171.90	710.0
173.00	760.0
174.00	72632.0
174.90	5270.0
176.00	69640.0
177.00	4876.0
178.00	197.0
190.90	342.0
193.10	520.0
207.00	400.0
249.10	211.0
250.00	151.0
265.00	522.0
281.10	2854.0
282.00	852.0
283.10	395.0
284.10	244.0

Instrument :
MSVOA_D
ClientSampleId :
BFB120

m/z	Abundance
35.90	225.0
37.00	2569.0
38.00	2482.0
39.10	994.0
40.00	1001.0
43.90	1349.0
45.10	637.0
46.00	194.0
47.00	676.0
47.90	556.0
49.10	2699.0
50.00	11537.0
51.10	3519.0
51.90	219.0
56.00	914.0
57.00	2145.0
60.00	688.0
61.10	3139.0
62.10	3849.0
63.00	2518.0
64.00	387.0
64.20	375.0
65.00	292.0
67.00	271.0
68.00	6207.0
69.10	6458.0
70.10	460.0
72.00	462.0
73.10	2819.0
74.00	10310.0
75.10	31792.0
76.10	3240.0
76.90	331.0
77.90	376.0
79.00	3204.0
80.00	910.0
81.00	3777.0
81.90	561.0
86.90	4016.0
88.00	3844.0
90.80	260.0
92.00	2186.0
93.00	2862.0
94.00	8456.0
95.10	62672.0
96.00	4880.0
96.90	182.0
103.90	316.0
105.80	157.0
115.80	208.0
117.00	643.0
117.70	248.0
118.80	416.0
126.00	298.0
127.80	280.0
129.90	485.0
132.80	225.0
140.90	830.0
142.90	1164.0
144.90	663.0
154.80	165.0
172.00	534.0
172.80	554.0
174.00	55688.0
175.00	4146.0
176.00	55672.0
176.90	3462.0
191.00	348.0
193.20	364.0
207.10	393.0
249.00	326.0
249.90	156.0
265.10	246.0
281.10	1870.0
282.10	584.0
283.30	323.0

Instrument :
MSVOA_D
ClientSampleId :
BFB120

m/z	Abundance
36.00	511.0
37.00	2154.0
38.10	2264.0
39.10	1269.0
40.00	912.0
44.10	1005.0
45.10	621.0
47.20	919.0
48.10	572.0
49.20	1780.0
50.10	11160.0
51.00	3990.0
52.10	153.0
55.00	168.0
56.00	918.0
57.10	1703.0
60.00	706.0
61.10	3461.0
62.10	3258.0
63.10	2844.0
64.00	344.0
65.10	388.0
66.70	260.0
68.00	7046.0
69.00	7370.0
70.00	568.0
72.00	530.0
73.00	3620.0
74.00	11885.0
75.10	34088.0
76.10	3173.0
77.10	331.0
78.10	327.0
79.00	3766.0
80.00	889.0
80.90	4113.0
82.00	871.0
86.00	162.0
87.10	4198.0
88.10	3989.0
90.90	246.0
92.00	2275.0
93.00	3345.0
94.10	8931.0
95.10	74376.0
96.00	4778.0
103.90	367.0
104.10	370.0
105.80	178.0
110.90	219.0
115.90	271.0
117.00	612.0
117.80	387.0
118.90	722.0
125.10	354.0
127.90	152.0
129.80	378.0
130.90	158.0
133.00	565.0
134.00	248.0
134.80	300.0
136.70	229.0
140.80	812.0
142.80	1413.0
144.90	495.0
147.90	237.0
152.80	210.0
154.80	279.0
170.90	201.0
171.90	555.0
173.00	632.0
174.00	70384.0
175.00	5018.0
176.00	66648.0
176.90	4601.0
190.80	401.0
192.70	181.0
193.10	304.0
207.10	450.0
249.10	278.0
265.00	339.0
281.10	3463.0
282.00	1186.0
283.10	717.0
284.10	189.0

Instrument :
MSVOA_D
ClientSampleId :
BFB120