

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050823\
 Data File : VD076037.D
 Acq On : 08 May 2023 14:04
 Operator : KP/SY
 Sample : 02591-29
 Misc : 4.51G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JREE3

Quant Time: May 09 02:37:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 06 00:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	490981	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	415843	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	133141	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	134608	12.497	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.000%
7) Chloroethane-d5	2.799	69	139823	15.025	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.080%
11) 1,1-Dichloroethene-d2	3.911	65	33314	13.416	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.680%
21) 2-Butanone-d5	6.981	46	50380	53.880	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.760%
24) Chloroform-d	7.569	84	238083	19.391	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.560%
26) 1,2-Dichloroethane-d4	8.234	65	108327	20.661	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.640%
32) Benzene-d6	8.199	84	418219	17.255	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.040%
36) 1,2-Dichloropropane-d6	9.210	67	147362	21.966	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.880%
41) Toluene-d8	10.269	98	355020	15.646	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	62.600%
43) trans-1,3-Dichloroprop...	10.528	79	49955	18.015	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.080%
47) 2-Hexanone-d5	10.881	63	41029	50.793	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.580%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	132339	25.779	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.120%
66) 1,2-Dichlorobenzene-d4	13.816	152	104774	22.248	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.000%
Target Compounds					Qvalue	
13) Acetone	4.016	43	8144	10.178	ug/L	87
16) Methylene chloride	4.793	84	28273	2.839	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050823\
Data File : VD076037.D
Acq On : 08 May 2023 14:04
Operator : KP/SY
Sample : 02591-29
Misc : 4.51G/10.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JREE3

Quant Time: May 09 02:37:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 06 00:21:26 2023
Response via : Initial Calibration

