

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD051519\
 Data File : VD062395.D
 Acq On : 15 May 2019 18:29
 Operator : FY/SY
 Sample : K2822-03
 Misc : 5.05µ/5mL/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-3

Quant Time: May 16 03:58:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	601	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.21	114	1237	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.37	117	825	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	194	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	10.41	98	791	31.96	ug/l	0.01
Spiked Amount	50.000		Recovery	=	63.92%	
62) 4-Bromofluorobenzene	13.56	95	255	25.66	ug/l	0.01
Spiked Amount	50.000		Recovery	=	51.32%	
Target Compounds						
					Qvalue	
18) Methyl Acetate	3.59	43	187	89.508	ug/l #	53
25) 2-Butanone	5.99	43	198	154.981	ug/l #	56
37) Ethyl Acetate	5.99	43	198	46.034	ug/l #	64
43) Isopropyl Acetate	8.93	43	213	34.918	ug/l #	43
59) 2-Hexanone	11.42	43	250	86.504	ug/l #	27
74) N-amyl acetate	13.09	43	293	96.101	ug/l #	49
84) 1,2,4-Trimethylbenzene	14.24	105	183	18.623	ug/l #	30
85) sec-Butylbenzene	14.24	105	183	15.164	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD051519\
Data File : VD062395.D
Acq On : 15 May 2019 18:29
Operator : FY/SY
Sample : K2822-03
Misc : 5.05g/5mL/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-3

Quant Time: May 16 03:58:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration

