

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051823\
 Data File : VD076139.D
 Acq On : 18 May 2023 14:32
 Operator : KP/SY
 Sample : 02816-01
 Misc : 6.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CB-22-K2-SO-3.5-051623

Quant Time: May 19 01:53:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	186654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	384785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	368497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	173178	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	137265	76.091	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.180%
35) Dibromofluoromethane	7.804	113	165543	62.241	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	124.480%
50) Toluene-d8	10.269	98	508165	49.012	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.020%
62) 4-Bromofluorobenzene	12.575	95	145400	50.484	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051823\
Data File : VD076139.D
Acq On : 18 May 2023 14:32
Operator : KP/SY
Sample : 02816-01
Misc : 6.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CB-22-K2-SO-3.5-051623

Quant Time: May 19 01:53:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

