

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
 Data File : VD076159.D
 Acq On : 19 May 2023 16:11
 Operator : KP/SY
 Sample : 02845-01
 Misc : 4.65G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1

Quant Time: May 22 01:30:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.863	168	8618	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.781	114	16734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	19454	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	6942	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	10066	120.854	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	241.700%#
35) Dibromofluoromethane	7.804	113	9931	85.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	171.720%#
50) Toluene-d8	10.275	98	19002	42.142	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.280%
62) 4-Bromofluorobenzene	12.569	95	6500	51.894	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
Data File : VD076159.D
Acq On : 19 May 2023 16:11
Operator : KP/SY
Sample : 02845-01
Misc : 4.65G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1

Quant Time: May 22 01:30:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

