

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
 Data File : VD076160.D
 Acq On : 19 May 2023 16:40
 Operator : KP/SY
 Sample : 02845-11
 Misc : 5.13G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP2

Quant Time: May 22 01:30:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	213453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	433027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	379875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	132382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	135374	65.621	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.240%
35) Dibromofluoromethane	7.804	113	156587	52.315	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.620%
50) Toluene-d8	10.269	98	465820	39.922	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.840%
62) 4-Bromofluorobenzene	12.575	95	133561	41.207	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.420%
Target Compounds						
16) Acetone	4.016	43	9042	22.029	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
Data File : VD076160.D
Acq On : 19 May 2023 16:40
Operator : KP/SY
Sample : 02845-11
Misc : 5.13G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP2

Quant Time: May 22 01:30:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

