

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
 Data File : VD076163.D
 Acq On : 19 May 2023 18:39
 Operator : KP/SY
 Sample : 02849-11
 Misc : 4.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP8

Quant Time: May 22 01:31:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	5350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	12224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	13451	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7196	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	6764	130.816	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	261.640%#
35) Dibromofluoromethane	7.799	113	6207	73.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	146.920%
50) Toluene-d8	10.269	98	13382	40.628	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	81.260%
62) 4-Bromofluorobenzene	12.575	95	8116	88.702	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	177.400%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051923\
Data File : VD076163.D
Acq On : 19 May 2023 18:39
Operator : KP/SY
Sample : 02849-11
Misc : 4.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP8

Quant Time: May 22 01:31:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

