

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
 Data File : VD076191.D
 Acq On : 23 May 2023 13:04
 Operator : KP/SY
 Sample : 02826-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-3-051723

Quant Time: May 23 15:02:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	166352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	324013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	290275	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	120255	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	96581	60.072	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.140%
35) Dibromofluoromethane	7.804	113	110648	49.404	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.269	98	390129	44.685	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.360%
62) 4-Bromofluorobenzene	12.575	95	98169	40.478	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
Data File : VD076191.D
Acq On : 23 May 2023 13:04
Operator : KP/SY
Sample : 02826-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-3-051723

Quant Time: May 23 15:02:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

