

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
 Data File : VD076199.D
 Acq On : 23 May 2023 18:00
 Operator : KP/SY
 Sample : 02864-05
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-359

Quant Time: May 24 06:42:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	34403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	47229	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	1400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	22956	69.041	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	138.080%	
35) Dibromofluoromethane	7.805	113	29771	91.195	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	182.380%#	
50) Toluene-d8	10.269	98	26108	20.515	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	41.040%#	
62) 4-Bromofluorobenzene	12.575	95	2120	5.997	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	12.000%#	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	21130	37.877	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
Data File : VD076199.D
Acq On : 23 May 2023 18:00
Operator : KP/SY
Sample : 02864-05
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-359

Quant Time: May 24 06:42:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

