

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
 Data File : VD076204.D
 Acq On : 23 May 2023 20:29
 Operator : KP/SY
 Sample : 02867-21
 Misc : 5.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP14

Quant Time: May 24 06:43:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	304686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	626220	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	552217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	224553	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	130248	44.231	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.460%
35) Dibromofluoromethane	7.804	113	171085	39.525	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.040%
50) Toluene-d8	10.269	98	508628	30.143	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	60.280%
62) 4-Bromofluorobenzene	12.575	95	186393	39.766	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
Data File : VD076204.D
Acq On : 23 May 2023 20:29
Operator : KP/SY
Sample : 02867-21
Misc : 5.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP14

Quant Time: May 24 06:43:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

