

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
 Data File : VD076205.D
 Acq On : 23 May 2023 20:59
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 24 06:43:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.451	0.380	15.7	70	0.00
3 P	Chloromethane	0.857	0.780	9.0	76	0.00
4 C	Vinyl Chloride	0.973	0.944	3.0#	78	0.00
5 T	Bromomethane	0.655	0.552	15.7	66	0.00
6 T	Chloroethane	0.654	0.650	0.6	79	0.00
7 T	Trichlorofluoromethane	0.824	0.716	13.1	73	0.00
8 T	Diethyl Ether	0.271	0.263	3.0	71	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.543	0.483	11.0	75	0.00
10 T	Methyl Iodide	0.565	0.550	2.7	67	0.00
11 T	Tert butyl alcohol	0.026	0.029	-11.5	85	0.00
12 CM	1,1-Dichloroethene	0.540	0.472	12.6#	69	0.00
13 T	Acrolein	0.063	0.047	25.4#	53	0.00
14 T	Allyl chloride	0.548	0.506	7.7	68	0.00
15 T	Acrylonitrile	0.112	0.126	-12.5	81	0.00
16 T	Acetone	0.096	0.082	14.6	60	0.00
17 T	Carbon Disulfide	1.620	1.278	21.1	64	0.00
18 T	Methyl Acetate	0.343	0.382	-11.4	81	0.00
19 T	Methyl tert-butyl Ether	1.058	1.070	-1.1	71	0.00
20 T	Methylene Chloride	1.029	0.688	33.1#	65	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.555	10.9	66	0.00
22 T	Diisopropyl ether	1.273	1.289	-1.3	70	0.00
23 T	Vinyl Acetate	0.501	0.525	-4.8	72	0.00
24 P	1,1-Dichloroethane	1.004	0.948	5.6	72	0.00
25 T	2-Butanone	0.121	0.127	-5.0	75	0.00
26 T	2,2-Dichloropropane	0.819	0.724	11.6	69	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.644	6.8	67	0.00
28 T	Bromochloromethane	0.335	0.336	-0.3	76	0.00
29 T	Tetrahydrofuran	0.070	0.081	-15.7	80	0.00
30 C	Chloroform	1.074	1.015	5.5#	72	0.00
31 T	Cyclohexane	0.740	0.664	10.3	73	0.00
32 T	1,1,1-Trichloroethane	0.856	0.778	9.1	71	0.00
33 S	1,2-Dichloroethane-d4	0.483	0.454	6.0	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.346	0.299	13.6	70	0.00
36 T	1,1-Dichloropropene	0.467	0.421	9.9	72	0.00
37 T	Ethyl Acetate	0.162	0.177	-9.3	80	0.00
38 T	Carbon Tetrachloride	0.422	0.364	13.7	72	0.00
39 T	Methylcyclohexane	0.537	0.468	12.8	73	0.00
40 TM	Benzene	1.501	1.348	10.2	70	0.00
41 T	Methacrylonitrile	0.089	0.092	-3.4	71	0.00
42 TM	1,2-Dichloroethane	0.355	0.331	6.8	73	0.00
43 T	Isopropyl Acetate	0.302	0.304	-0.7	78	0.00
44 TM	Trichloroethene	0.423	0.353	16.5	67	0.00
45 C	1,2-Dichloropropane	0.370	0.339	8.4#	72	0.00
46 T	Dibromomethane	0.207	0.192	7.2	73	0.00
47 T	Bromodichloromethane	0.499	0.452	9.4	71	0.00
48 T	Methyl methacrylate	0.134	0.134	0.0	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
 Data File : VD076205.D
 Acq On : 23 May 2023 20:59
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 24 06:43:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	86	0.00
50 S	Toluene-d8	1.347	1.186	12.0	70	0.00
51 T	4-Methyl-2-Pentanone	0.161	0.176	-9.3	81	0.00
52 CM	Toluene	0.937	0.862	8.0#	71	0.00
53 T	t-1,3-Dichloropropene	0.449	0.422	6.0	71	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.506	7.7	70	0.00
55 T	1,1,2-Trichloroethane	0.298	0.288	3.4	75	0.00
56 T	Ethyl methacrylate	0.299	0.310	-3.7	74	0.00
57 T	1,3-Dichloropropane	0.467	0.450	3.6	74	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.153	-7.0	79	0.00
59 T	2-Hexanone	0.113	0.121	-7.1	76	0.00
60 T	Dibromochloromethane	0.335	0.314	6.3	73	0.00
61 T	1,2-Dibromoethane	0.265	0.262	1.1	77	0.00
62 S	4-Bromofluorobenzene	0.374	0.333	11.0	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.384	0.302	21.4	66	0.00
65 PM	Chlorobenzene	1.138	0.987	13.3	71	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.344	12.2	73	0.00
67 C	Ethyl Benzene	1.819	1.653	9.1#	73	0.00
68 T	m/p-Xylenes	0.743	0.689	7.3	74	0.00
69 T	o-Xylene	0.678	0.631	6.9	72	0.00
70 T	Styrene	1.176	1.129	4.0	75	0.00
71 P	Bromoform	0.220	0.203	7.7	75	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.308	2.939	11.2	75	0.00
74 T	N-amyl acetate	0.589	0.565	4.1	77	0.00
75 P	1,1,2,2-Tetrachloroethane	0.711	0.685	3.7	84	0.00
76 T	1,2,3-Trichloropropane	0.438	0.345	21.2	77	0.00
77 T	Bromobenzene	0.887	0.740	16.6	70	0.00
78 T	n-propylbenzene	4.117	3.705	10.0	76	0.00
79 T	2-Chlorotoluene	2.279	1.993	12.5	74	0.00
80 T	1,3,5-Trimethylbenzene	2.742	2.451	10.6	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.209	0.203	2.9	81	0.00
82 T	4-Chlorotoluene	2.424	2.099	13.4	73	0.00
83 T	tert-Butylbenzene	2.317	2.100	9.4	77	0.00
84 T	1,2,4-Trimethylbenzene	2.731	2.475	9.4	74	0.00
85 T	sec-Butylbenzene	3.611	3.295	8.8	79	0.00
86 T	p-Isopropyltoluene	2.955	2.671	9.6	76	0.00
87 T	1,3-Dichlorobenzene	1.792	1.550	13.5	74	0.00
88 T	1,4-Dichlorobenzene	1.775	1.528	13.9	75	0.00
89 T	n-Butylbenzene	2.785	2.438	12.5	75	0.00
90 T	Hexachloroethane	0.608	0.514	15.5	77	0.00
91 T	1,2-Dichlorobenzene	1.573	1.365	13.2	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.092	0.086	6.5	79	0.00
93 T	1,2,4-Trichlorobenzene	0.915	0.711	22.3	63	0.00
94 T	Hexachlorobutadiene	0.467	0.338	27.6#	64	0.00
95 T	Naphthalene	1.700	1.589	6.5	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052323\
Data File : VD076205.D
Acq On : 23 May 2023 20:59
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 24 06:43:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.832	0.676	18.7	65	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6