

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
 Data File : VD076219.D
 Acq On : 24 May 2023 17:39
 Operator : KP/SY
 Sample : 02892-13
 Misc : 5.55G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-29

Quant Time: May 25 06:02:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	260200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	550362	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	377637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	74667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	132355	52.631	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.260%
35) Dibromofluoromethane	7.804	113	188148	49.458	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.920%
50) Toluene-d8	10.269	98	440938	29.733	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	59.460%
62) 4-Bromofluorobenzene	12.575	95	91175	22.133	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	44.260%
Target Compounds						
20) Methylene Chloride	4.799	84	82400	14.098	ug/l #	80
64) Tetrachloroethene	10.810	164	10346	3.563	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
Data File : VD076219.D
Acq On : 24 May 2023 17:39
Operator : KP/SY
Sample : 02892-13
Misc : 5.55G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-29

Quant Time: May 25 06:02:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

