

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
 Data File : VD076220.D
 Acq On : 24 May 2023 18:09
 Operator : KP/SY
 Sample : 02891-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 290-328

Quant Time: May 25 06:02:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	237513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	497472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	340843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	62423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	158436	69.020	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.040%	
35) Dibromofluoromethane	7.805	113	220281	64.061	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.120%	
50) Toluene-d8	10.269	98	488508	36.443	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 72.880%	
62) 4-Bromofluorobenzene	12.575	95	87777	23.573	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 47.140%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	64608	10.629	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
Data File : VD076220.D
Acq On : 24 May 2023 18:09
Operator : KP/SY
Sample : 02891-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
290-328

Quant Time: May 25 06:02:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

