

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
 Data File : VD076224.D
 Acq On : 24 May 2023 20:10
 Operator : KP/SY
 Sample : 02867-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP16

Quant Time: May 25 06:03:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	370894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	774579	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	696199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	278964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	168064	46.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.780%
35) Dibromofluoromethane	7.804	113	228345	42.649	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.300%
50) Toluene-d8	10.269	98	680127	32.587	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	65.180%
62) 4-Bromofluorobenzene	12.575	95	234274	40.408	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.820%
Target Compounds						
16) Acetone	4.028	43	17734	24.865	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052423\
Data File : VD076224.D
Acq On : 24 May 2023 20:10
Operator : KP/SY
Sample : 02867-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP16

Quant Time: May 25 06:03:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

