

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073362.D
 Acq On : 25 May 2022 15:56
 Operator : VA/SY
 Sample : N2842-23
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P025-SS001-0612-01

Quant Time: May 26 02:25:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	107029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	187772	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	214544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	101356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	84481	71.423	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	142.840%	
35) Dibromofluoromethane	7.908	113	74202	59.886	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	119.780%	
50) Toluene-d8	10.332	98	233691	51.165	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	102.320%	
62) 4-Bromofluorobenzene	12.620	95	96465	60.182	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	120.360%	
Target Compounds						
11) Tert butyl alcohol	5.191	59	11219	61.134	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073362.D
Acq On : 25 May 2022 15:56
Operator : VA/SY
Sample : N2842-23
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P025-SS001-0612-01

Quant Time: May 26 02:25:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

